UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/552,030	09/12/2006	Gianluca Gazza	82062-0169	6978
<sup>24633</sup> HOGAN & HA	7590 12/18/200 RTSON LLP	EXAMINER		
	DLUMBIA SQUARE		BERRIOS, JENNIFER ANN	
555 THIRTEENTH STREET, N.W. WASHINGTON, DC 20004			ART UNIT	PAPER NUMBER
			4121	
			NOTIFICATION DATE	DELIVERY MODE
			12/18/2008	ELECTRONIC

## Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Notice of the Office communication was sent electronically on above-indicated "Notification Date" to the following e-mail address(es):

dcptopatent@hhlaw.com rogruwell@hhlaw.com

	Application No.	Applicant(s)				
	10/552,030	GAZZA, GIANLUCA				
Office Action Summary	Examiner	Art Unit				
	Jennifer A. Berríos	4121				
The MAILING DATE of this communication app Period for Reply	ears on the cover sheet with the c	orrespondence address				
A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 1 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.  - Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.  - If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.  - Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).						
Status						
1) Responsive to communication(s) filed on						
	action is non-final.					
·=	/ <del></del>					
	closed in accordance with the practice under <i>Ex parte Quayle</i> , 1935 C.D. 11, 453 O.G. 213.					
Disposition of Claims						
4)⊠ Claim(s) <u>1-42</u> is/are pending in the application.						
4a) Of the above claim(s) is/are withdrawn from consideration.						
5) Claim(s) is/are allowed.						
6) Claim(s) is/are rejected.						
7) Claim(s) is/are objected to.						
8) Claim(s) <u>1-42</u> are subject to restriction and/or e	lection requirement.					
Application Papers						
9) The specification is objected to by the Examiner.						
10) The drawing(s) filed on is/are: a) accepted or b) objected to by the Examiner.						
Applicant may not request that any objection to the o	• , ,	* *				
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).  11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.						
TT) The bath or declaration is objected to by the Exa	aminer, Note the attached Office	Action of form PTO-152.				
Priority under 35 U.S.C. § 119						
<ul> <li>12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).</li> <li>a) All b) Some * c) None of:</li> <li>1. Certified copies of the priority documents have been received.</li> <li>2. Certified copies of the priority documents have been received in Application No</li> <li>3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).</li> <li>* See the attached detailed Office action for a list of the certified copies not received.</li> </ul>						
Attachment(s)  1) ☑ Notice of References Cited (PTO-892)	4) ☐ Interview Summary	(PTO-413)				
2) Notice of Traftsperson's Patent Drawing Review (PTO-948)	Paper No(s)/Mail Da	ate				
3) Information Disclosure Statement(s) (PTO/SB/08)	5) Notice of Informal P.	atent Application				
Paper No(s)/Mail Date 6) U Other:						

## **DETAILED ACTION**

## Election/Restrictions

1. Restriction is required under 35 U.S.C. 121 and 372.

This application contains the following inventions or groups of inventions which are not so linked as to form a single general inventive concept under PCT Rule 13.1.

In accordance with 37 CFR 1.499, applicant is required, in reply to this action, to elect a single invention to which the claims must be restricted.

Group I, claims 1-3, drawn to a stent comprising a coating based on polymer of hyaluronic acid.

Group II, claims 33-40, drawn to a process for obtaining a coated stent.

Group II, claims 41-42, drawn to a use of a hyaluronic ester for the preparation of a coating for a stent.

As set forth in Rule 13.1 of the Patent Cooperation Treaty (PCT), "the international application shall relate to one invention only or to a group of inventions so linked as to form a single general inventive concept." Moreover, as stated in PCT Rule 13.2, "where a group of inventions is claimed in one and the same international application, the requirement of unity of invention referred to in Rule 13.1 shall be fulfilled only when there is a technical relationship among those inventions involving one or more of the same or corresponding special technical features." Furthermore, Rule 13.2 defines "special technical features" as "those technical features that define a contribution which each of the claimed inventions, considered as a whole, makes over the prior art."

Application/Control Number: 10/552,030

Art Unit: 4121

The inventions listed as Groups I, II and III do not relate to a single general inventive concept under PCT Rule 13.1 because, under PCT Rule 13.2, they lack the same or corresponding special technical features for the following reasons:

Page 3

2. The special technical feature of Group I is a stent comprising a coating based on a polymer of hyaluronic acid. The stent comprising a coating based on a polymer of hyaluronic acid of claim 1 does not present a contribution over the prior art. US Patent 6027741 (Cialdi et Al, Column 2, lines 11-15), provided by Applicant, discloses a coated biomedical object or device having a coating of sulfated polysaccharide, wherein the polysaccharide is a hyaluronic acid, hyluronate ester, or a salt thereof. This art fails to teach what constitutes a biomedical object or device. A secondary reference, article Biomedical Implants and Devices: Assessment of Magnetic Field Interactions with a 3.0-Tesla MR System (Shellock, Column 1, lines 4-6) discloses that biomedical devices and objects consist of stents, fasteners, staples and more; thus teaching what the prior art fails to teach. These two references are being combined because it is well known to one of ordinary skill in the art that a stent is a biomedical device. Therefore, Claim1 is rejected under 35 U.S.C. 103(a) as being unpatentable over the prior art.

As disclosed in Cialdi Gloria et Al (US 6027741) (Column 2, Lines 11-15), the stent comprising a coating based on a polymer of hyaluronic acid of instant claim 1 is not novel or does not involve an inventive step. As such, Group I does not share a special technical feature with the instant claims of Group II and II. Therefore, the claims are not so linked within the meaning of PCT Rule 13.2 so as to form a single inventive concept, and unity between Groups I-III is broken.

Application/Control Number: 10/552,030 Page 4

Art Unit: 4121

## Election of Species

3. This application contains claims directed to more than one species of the generic invention. These species are deemed to lack unity of invention because they are not so linked as to form a single general inventive concept under PCT Rule 13.1.

The species are as follows:

If Group I is elected, EACH of the following species elections are required:

a) One specific alcohol, with claims 2-3 and 6-7 reading upon this species.

Applicant is required, in reply to this action, to elect a single species to which the claims shall be restricted if no generic claim is finally held to be allowable. Specifically, as to claims 2-3, <u>Applicant is required</u> to select a specific alcohol or a defined plurality of alcohols. As to claim 2, Applicant is required to elect a single specific alcohol.

Upon the allowance of a generic claim, applicant will be entitled to consideration of claims to additional species which are written in dependent form or otherwise include all the limitations of an allowed generic claim as provided by 37 CFR 1.141. If claims are added after the election, applicant must indicate which are readable upon the elected species. MPEP § 809.02(a).

4. The claims are deemed to correspond to the species listed above in the following manner:

Claims 2-3 and 6-7 as to Group I, for the election of a specific alcohol.

The following claim(s) are generic: 1-2, 4-5 and 8-32 for Group I.

The species listed above do not relate to a single general inventive concept under PCT Rule 13.1 because, under PCT Rule 13.2, the species lack the same or corresponding special technical features for the following reasons:

To those of ordinary skill in the art, it is well known that ethyl alcohol has been used prior to the Applicant's invention.

There is an examination and search burden for these patentably distinct species due to their mutually exclusive characteristics. The species require a different field of search (e.g., searching different classes/subclasses or electronic resources, or employing different search queries); and/or the prior art applicable to one species would not likely be applicable to another species; and/or the species are likely to raise different non-prior art issues under 35 U.S.C. 101 and/or 35 U.S.C. 112, first paragraph.

5. Applicant is advised that the reply to this requirement to be complete must include (i) an election of a species or invention to be examined even though the requirement may be traversed (37 CFR 1.143) and (ii) identification of the claims encompassing the elected invention.

The election of an invention or species may be made with or without traverse. To preserve a right to petition, the election must be made with traverse. If the reply does not distinctly and specifically point out supposed errors in the restriction requirement, the election shall be treated as an election without traverse.

6. The examiner has required restriction between product and process claims.

Where applicant elects claims directed to the product, and the product claims are subsequently found allowable, withdrawn process claims that depend from or otherwise require all the limitations of the allowable product claim will be considered for rejoinder.

All claims directed to a nonelected process invention must require all the limitations of an allowable product claim for that process invention to be rejoined.

Application/Control Number: 10/552,030

Art Unit: 4121

In the event of rejoinder, the requirement for restriction between the product claims and the rejoined process claims will be withdrawn, and the rejoined process claims will be fully examined for patentability in accordance with 37 CFR 1.104. Thus, to be allowable, the rejoined claims must meet all criteria for patentability including the requirements of 35 U.S.C. 101, 102, 103 and 112. Until all claims to the elected product are found allowable, an otherwise proper restriction requirement between product claims and process claims may be maintained. Withdrawn process claims that are not commensurate in scope with an allowable product claim will not be rejoined. See MPEP § 821.04(b). Additionally, in order to retain the right to rejoinder in accordance with the above policy, applicant is advised that the process claims should be amended during prosecution to require the limitations of the product claims. **Failure to do so may result in a loss of the right to rejoinder**. Further, note that the prohibition against double patenting rejections of 35 U.S.C. 121 does not apply where the restriction requirement

Page 6

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Jennifer A. Berríos whose telephone number is (571)270-7679. The examiner can normally be reached on M-F 8:00am-5:00pm (EST).

is withdrawn by the examiner before the patent issues. See MPEP § 804.01.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Patrick Nolan can be reached on (571) 272-0847. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Application/Control Number: 10/552,030 Page 7

Art Unit: 4121

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

JAB

/Patrick J. Nolan/ Supervisory Patent Examiner, Art Unit 4121